

21—85.53(215A) Rejecting devices. Moisture-measuring devices may be rejected for any of the following reasons:

85.53(1) The moisture-measuring device tested is found to be out of tolerance with the measuring device used by the department by one of the inspectors so assigned by more than 0.7 percent on grain moisture content.

85.53(2) The person does not have available the latest charts for type of device being used.

85.53(3) The person does not have available the proper scale or scales and thermometers for use with the type of device being used.

85.53(4) The moisture-measuring device is not free from excessive dirt, debris, cracked glass or is not kept in good operational condition at all times.

This rule is intended to implement Iowa Code section 215A.6.